



DECLARATION OF CONFORMITY

ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

EU Representative

SUNGO Europe B.V.
Olympisch Stadion 24, 1076DE
Amsterdam, Netherlands
SRN: NL-AR-000000247

Conformity Assessment

Conformity Assessment Procedure
Annex II+III of Regulation (EU) 2017/745

Applicable Standards
EN ISO 14971: 2019
EN ISO 15223-1: 2016
EN 1041:2008+A1:2013
EN 60601-1-2:2015
EN 60601-1:2006+A12:2014

Remark

The declaration of conformity is valid in connection with the release technical document CE/MDR-JF-01.

All the supporting documentation is retained at the premises of the manufacturer.

The Declaration of Conformity is exclusively under the sole responsibility of the manufacturer.

Manufacturer

Name: Shenzhen Beacon Display Technology Co., Ltd.

Address: 15F, Building 6, Hengda Shishang Huigu(East) , Fulong Road, Henglang Community, Dalang street, Longhua, Shenzhen, Guangdong, China

SRN:CN-MF-000008774

Product Information

Name: LCD Monitor

Model: See annex 1


GMDN: 36612/36615/56654

Basic UDI-DI: 697002782001FH

Classification: Class I, According to Rule 1, Annex VIII, Regulation (EU) 2017/745

Declaration

We herewith declare that the above-mentioned products meet the requirements of Medical Device Regulation (EU) 2017/745 and the applicable standards above.

Signature: 

Position: GM




Place: Guangdong/China

Annex 1

Product Name	Model	GMDN	Basic UDI-DI
LCD Monitor	C22S+、C22SP+、C24S+、G22S+、G22SP+、G23S+、G23SP+、C32S+、C32SP+、G32S+、G32SP+、C53S+、C53SP+、G52S+、G52SP+、G53S+、G53SP+、C43W+、C44W+、C61W+、C81W+、C82W+、C83W+、C85W+、C86W+、C14ST、C14S、G11S、C22W、C22WT、C23W、C23WT、C24W、C24WT、M553T、M652T、M862T、M982T、E150、E190、E2421、E2422、E2721、S2421P、S271P、S272P、S273P、S2785P、S3281P、S3285P、S3180P、S421、S551、S5583P、S5580P、PV26、PV27	36612/36615 /56654	/

Signature:



Position: GM



Date: 2021-05-12

Place: Guangdong/China